



**WHO International Standard**  
**Golimumab**  
**NIBSC code: 22/116**  
**Instructions for use**  
**(Version 1.0, Dated 05/04/2024)**

§

## 1. INTENDED USE

The World Health Organization (WHO) Expert Committee on Biological Standardization (ECBS) recognised the need for a reference standard to evaluate the performance of in vitro biological assays for golimumab.

The International Standard 22/116 is intended to support the calibration, characterisation and validation of assays used for assessing golimumab and to support the establishment of in-house standards.

The standard was assessed in an international collaborative study (described in section 3), for in vitro biological activities of golimumab. The standard was also assessed for use in assays for therapeutic drug monitoring.

## 2. CAUTION

**This preparation is not for administration to humans or animals in the human food chain.**

The preparation contains material of human origin, and either the final product or the source materials, from which it is derived, have been tested and found negative for HBsAg, anti-HIV and HCV RNA. As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols. Care should be exercised in opening ampoules or vials, to avoid cuts.

## 3. UNITAGE

**Bioactivity:** The preparation has been assigned the following arbitrary unitage per ampoule:

500 international units (IU)\* of tumor necrosis factor (TNF) neutralising activity.  
500 IU of TNF binding activity.  
500 IU of Fcγ<sub>3</sub> binding activity.  
500 IU of ADCC activity.

\*These units are independent of the amount of TNF used in various bioassays. For details regarding neutralising activity in terms of the 3rd IS for TNF (coded 12/154), see report referenced in section 9.

It should be noted that the neutralising activity may vary according to the assay format. Therefore, a relationship between the unitage of the WHO IS coded 22/116 and the activity assigned to in-house standards in the assay system in routine use should be established.

Users should also note that the biological activity of TNF is likely to vary between different suppliers and this should be controlled by use of an appropriate standard (e.g. WHO IS).

The golimumab IS was tested using assays established in-house in a multi-centre collaborative study involving 16 laboratories in 11 countries. Participants reported results for a range of assays e.g., cytotoxicity, binding (TNF & CD16) and antibody-dependant cellular cytotoxicity (see reference in section 9).

**Therapeutic drug monitoring (TDM):** The use of an assumed mass content of 50 micrograms of golimumab per ampoule is recommended only for TDM assays (see section 4). The suitability of the golimumab IS for TDM was assessed in a range of binding assays in a study involving 8 laboratories in 6 countries (see reference in section 9).

**Note:** It should be noted that the unitage or mass content of the standard should not be used to define the specific activity of golimumab products for regulatory purposes nor to describe product labelling or dosage requirements. Furthermore, the standard and its unitage are not intended to serve any regulatory role in defining biosimilarity, and should not be inferred as serving this purpose.

## 4. CONTENTS

Country of origin of biological material: Ireland.

Each ampoule contains the residue after freeze-drying of 1.0 ml of a solution containing:  
Golimumab, approximately 50 micrograms  
25mM tri-sodium citrate dihydrate  
150mM sodium chloride  
1.0% human serum albumin

## 5. STORAGE

Unopened ampoules should be stored at -20°C. For economy of use, it is recommended that the solution be sub divided into aliquots and stored at -40°C or below. Avoid repeated thawing/freezing.

Please note because of the inherent stability of lyophilized material, NIBSC may ship these materials at ambient temperature.

## 6. DIRECTIONS FOR OPENING

DIN ampoules have an 'easy-open' coloured stress point, where the narrow ampoule stem joins the wider ampoule body. Various types of ampoule breaker are available commercially. To open the ampoule, tap the ampoule gently to collect material at the bottom (labelled) end and follow manufactures instructions provided with the ampoule breaker.

## 7. USE OF MATERIAL

**No attempt should be made to weigh out any portion of the freeze-dried material prior to reconstitution**

Dissolve the total contents of the ampoule in 1.0ml of sterile distilled water. Use carrier protein where extensive dilution is required. Users should note that in rare instances interference due to excipients may occur if the IS is used at high concentrations ( ≥ 10µg/ml).

## 8. STABILITY

Reference materials are held at NIBSC within assured, temperature-controlled storage facilities. Reference Materials should be stored on receipt as indicated on the label.



NIBSC follows the policy of WHO with respect to its reference materials.

## 9. REFERENCES

This standard was produced under WHO guidelines cited in the WHO Technical Reports Series, No. 932, 2006, Annex 2.

Report on a Collaborative Study for Proposed 1<sup>st</sup> International Standard for Golimumab WHO/BS/2024.2467

## 10. ACKNOWLEDGEMENTS

We are thankful to Janssen Pharmaceuticals R&D LLC, USA, for their generous donation of the golimumab material used in the collaborative study. We are grateful to all participants of the collaborative study for their contribution in evaluating the candidate preparations.

## 11. FURTHER INFORMATION

Further information can be obtained as follows;

This material: [enquiries@nibsc.org](mailto:enquiries@nibsc.org)

WHO Biological Standards:

<http://www.who.int/biologicals/en/>

JCTLM Higher order reference materials:

<http://www.bipm.org/en/committees/jc/jctlm/>

Derivation of International Units:

[http://www.nibsc.org/standardisation/international\\_standards.aspx](http://www.nibsc.org/standardisation/international_standards.aspx)

Ordering standards from NIBSC:

<http://www.nibsc.org/products/ordering.aspx>

NIBSC Terms & Conditions:

[http://www.nibsc.org/terms\\_and\\_conditions.aspx](http://www.nibsc.org/terms_and_conditions.aspx)

## 12. CUSTOMER FEEDBACK

Customers are encouraged to provide feedback on the suitability or use of the material provided or other aspects of our service. Please send any comments to [enquiries@nibsc.org](mailto:enquiries@nibsc.org)

## 13. CITATION

In all publications, including data sheets, in which this material is referenced, it is important that the preparation's title, its status, the NIBSC code number, and the name and address of NIBSC are cited and cited correctly.

## 14. MATERIAL SAFETY SHEET

Classification in accordance with Directive 2000/54/EC, Regulation (EC) No 1272/2008: Not applicable or not classified

Physical and Chemical properties	
Physical appearance: Freeze dried powder	Corrosive: No
Stable: Yes	Oxidising: No
Hygroscopic: No	Irritant: No
Flammable: No	Handling: See caution, Section 2
Other (specify):	Contains material of human origin
Toxicological properties	
Effects of inhalation:	Not established, avoid inhalation
Effects of ingestion:	Not established, avoid ingestion
Effects of skin absorption:	Not established, avoid contact with skin

Suggested First Aid	
Inhalation:	Seek medical advice
Ingestion:	Seek medical advice
Contact with eyes:	Wash with copious amounts of water. Seek medical advice
Contact with skin:	Wash thoroughly with water.
Action on Spillage and Method of Disposal	
Spillage of ampoule contents should be taken up with absorbent material wetted with an appropriate disinfectant. Rinse area with an appropriate disinfectant followed by water. Absorbent materials used to treat spillage should be treated as biological waste.	

## 15. LIABILITY AND LOSS

In the event that this document is translated into another language, the English language version shall prevail in the event of any inconsistencies between the documents.

Unless expressly stated otherwise by NIBSC, NIBSC's Standard Terms and Conditions for the Supply of Materials (available at [http://www.nibsc.org/About\\_Us/Terms\\_and\\_Conditions.aspx](http://www.nibsc.org/About_Us/Terms_and_Conditions.aspx) or upon request by the Recipient) ("Conditions") apply to the exclusion of all other terms and are hereby incorporated into this document by reference. The Recipient's attention is drawn in particular to the provisions of clause 11 of the Conditions.

## 16. INFORMATION FOR CUSTOMS USE ONLY

<b>Country of origin for customs purposes*:</b> United Kingdom * Defined as the country where the goods have been produced and/or sufficiently processed to be classed as originating from the country of supply, for example a change of state such as freeze-drying.
<b>Net weight:</b> 4.6g
<b>Toxicity Statement:</b> Toxicity not assessed
<b>Veterinary certificate or other statement</b> if applicable.
<b>Attached:</b> No

## 17. CERTIFICATE OF ANALYSIS

NIBSC does not provide a Certificate of Analysis for WHO Biological Reference Materials because they are internationally recognised primary reference materials fully described in the instructions for use. The reference materials are established according to the WHO Recommendations for the preparation, characterization and establishment of international and other biological reference standards [http://www.who.int/bloodproducts/publications/TRS932Annex2\\_1nter\\_biorefstandardsrev2004.pdf](http://www.who.int/bloodproducts/publications/TRS932Annex2_1nter_biorefstandardsrev2004.pdf) (revised 2004). They are officially endorsed by the WHO Expert Committee on Biological Standardization (ECBS) based on the report of the international collaborative study which established their suitability for the intended use.